

**WAMA: Water Activity and Moisture Analysis [WI-10-25]**

Analyst: KAR

Test Date: 11/18/2021

**99402-MA**Water Activity (Aw): **0.969**Moisture Content **69.5%**

The water activity was recorded with a dew point moisture analyzer. Water activity, Aw, is the partial vapor pressure of water in a substance divided by the partial vapor pressure of pure water at the same temperature. Moisture content was interpolated using an isotherm specific to a cannabis matrix.

**PST: Pesticide Analysis [WI-10-11]**

Analyst: CJR

Test Date: 11/15/2021

The client sample was analyzed for pesticides using Liquid Chromatography with Mass Spectrometric detection (LC/MS/MS). The method used for sample prep was based on the European method for pesticide analysis (EN 15662).

**99402-PST**

Analyte	CAS	Result	Units	LLD	Limits (ppb)	Status
Abamectin	71751-41-2	ND	ppb	0.20	10	PASS
Azoxystrobin	131860-33-8	ND	ppb	0.10	100	PASS
Bifenazate	149877-41-8	ND	ppb	0.10	100	PASS
Bifenthrin	82657-04-3	ND	ppb	0.20	3000	PASS
Cyfluthrin	68359-37-5	ND	ppb	0.50	2000	PASS
Dichlorvos	62-73-7	ND	ppb	3.00	10	PASS
Etoxazole	153233-91-1	ND	ppb	0.10	100	PASS
Fenoxycarb	72490-01-8	ND	ppb	0.10	10	PASS
Imazalil	35554-44-0	ND	ppb	0.10	10	PASS
Imidacloprid	138261-41-3	ND	ppb	0.10	5000	PASS
Myclobutanil	88671-89-0	ND	ppb	0.10	100	PASS
Paclobutrazol	76738-62-0	ND	ppb	0.10	10	PASS
Piperonyl butoxide	51-03-6	ND	ppb	0.10	3000	PASS
Pyrethrin	8003-34-7	ND	ppb	0.10	10	PASS
Spinosad	168316-95-8	ND	ppb	0.10	10	PASS
Spiromesifen	283594-90-1	ND	ppb	0.10	100	PASS
Spirotetramat	203313-25-1	ND	ppb	0.10	100	PASS
Trifloxystrobin	141517-21-7	ND	ppb	0.10	100	PASS

\* Testing limits established by the Massachusetts Department of Public Health, Protocol for Sampling and Analysis of Finished Medical Marijuana Products and Marijuana-Infused Products for Massachusetts Registered Medical Marijuana Dispensaries, Exhibit 5. ND indicates "none detected" above the lower limit of detection (LLD). Analytes marked with (\*) indicate analytes for which no recovery was observed for a pre-spiked matrix sample due to matrix interference.